

CENTRAL MEDICAL SERVICES SOCIETY  
Ministry of Health & Family Welfare  
(Autonomous Body under MoHFW, Govt. of India)  
2<sup>nd</sup> Floor, Vishwa Yuvak Kendra, Pandit Uma Shankar Dikshit Road,  
Chanakyapuri, New Delhi-110021

Date: - 12.06.2024

CORRIGENDUM-4

Tender Reference No: **CMSS/PROC/2024-25/NHM/008, Dated 13.05.2024**

Central Medical Services Society, an autonomous body under Ministry of Health and Family welfare, Government of India, had invited online tenders on Central Public Procurement Portal (CPPP), from eligible and qualified tenderers for **procurement of Point of Care Test Kits under National Sickle Cell Anemia Elimination Mission** As per the representation received from the vendor, the reply as received from Program division and CDSCO is detailed in the table below:

Details of item	UOM	Representation made by the vendor	Reply to the representation made by the firm
Rapid Diagnostic Test Kits for Sickle Cell Disease Diagnostic	Kits	Firstly, we would like to address the requirement stated in the tender for a class C specified license. Upon reviewing and analyzing the eligibility criteria, it has come to our attention that class B licenses are deemed eligible for participation as valid licenses issued by CDSCO. Therefore, we kindly request that the tender be amended to reflect the eligibility of class B licenses for participation.	The matter was referred to CDSCO and as per their response, Class-C Manufacturing License is mandatory. Hence, No Change.
		Secondly, there appears to be a discrepancy in the tender amendment regarding the sample volume required for our product. The amendment mentions the use of a graduated dropper with a 0-10ul capacity, whereas our product specifically requires a sample volume of 5ul and marked for appropriate sample. It is important to note that our product has been approved by the Indian Council of Medical Research (ICMR) and licensed by the Central Drugs Standard Control Organization (CDSCO), ensuring its compliance with regulatory standards. Hence this dropper graduation point should be deleted as it should depend exclusively on different manufacturers and their testing protocols.	<b>'Acceptable'</b>

**Note: - Apart from above, all other terms & conditions of tender document shall remain unchanged.**

